



LABGUARD 2

Installation report

UKC Ljubljana

Installation report

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
	c877ce050000	False	20	19	19	00:45:00	25	25	00:45:00	°C		
	24733f050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	dfe3c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	ae6fbf050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	eb65c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	67ac7d050000	True	25	15	15	00:45:00	35	35	00:45:00	°C		
	31b97d050000	False	-20	-40	-40	00:15:00	30	30	00:15:00	°C		
	f1f74c030000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	65147d050000	False	23	19	19	00:45:00	25	25	00:45:00	°C		
	d82028060000	True	5	2	2	00:15:00	8	8	00:15:00	°C		
	2a04ce040000	True	4	2	2	00:30:00	6	6	00:30:00	°C		
	037d1f050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	76c21f050000	False	5	2	2	00:45:00	8	8	00:45:00	°C		
	0ff61f050000	True	-30	-40	-40	00:45:00	-18	-18	00:45:00	°C		
	cf7ecb040000	False	17	14	14	00:45:00	24	24	00:45:00	°C		

Installation report

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
	b0601e050000	False	3	-6	-6	00:45:00	5	5	00:45:00	°C		
	sns0_10294c3001cd	False	-60	-85	-85	00:45:00	-55	-55	00:45:00	°C		
	sns_0_0b1e09300217	False	1	0.5	0.5	00:45:00	1.5	1.5	00:45:00	°C		
	9444cb040000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	sns_0_11234d3000f5	False	50	10	10	01:00:00	80	80	01:00:00	%rH		
	e7c1ca040000	False	20	10	10	00:45:00	30	30	00:45:00	°C		
	4699c0050000	False	-30	-35	-35	00:45:00	-25	-25	00:45:00	°C		
	sns_0_10234d30008f	True	50	35	35	00:45:00	65	65	00:45:00	%rH		
	2841cb040000	False	20	15	15	00:45:00	25	25	00:45:00	°C		
	sns_0_11234d3000b1	True	50	35	35	00:45:00	65	65	00:45:00	%rH		
	1e22cb040000	False	20	15	15	00:45:00	25	25	00:45:00	°C		
	sns_0_10234d3000ad	False	50	15	15	00:45:00	75	75	00:45:00	%rH		
	f23bce040000	False	20	15	15	00:45:00	25	25	00:45:00	°C		
	a121c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	63f63e050000	True	-20	-25	-25	00:45:00	-13	-13	00:45:00	°C		
	c893c0050000	False	-20	-25	-25	00:45:00	-15	-15	00:45:00	°C		
	4dbb7c050000	False	-20	-25	-25	00:45:00	-15	-15	00:45:00	°C		

Installation report

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
	3812cf050000	False	23	19	19	00:45:00	25	25	00:45:00	°C		
	1db9cc030000	True	-70	-80	-80	01:00:00	-57	-57	01:00:00	°C		
	59ff0c040000	False	-80	-90	-90	01:00:00	-70	-70	01:00:00	°C		
	c2eb2f040000	True	-80	-90	-90	01:00:00	-70	-70	01:00:00	°C		
	121ebc030000	True	-80	-90	-90	00:45:00	-70	-70	00:45:00	°C		
	8c71cc030000	True	-70	-85	-85	01:00:00	-60	-60	01:00:00	°C		
	ed7fc0050000	False	23	19	19	00:45:00	25	25	00:45:00	°C		
	12263e050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	2781bf050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	06af3e050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	eaba3e050000	False	5	2	2	00:45:00	8	8	00:45:00	°C		
	b177c0050000	True	5	2	2	00:45:00	12	12	00:45:00	°C		
	43403f050000	False	25	20	20	00:45:00	30	30	00:45:00	°C		
	2ed2be050000	False	5	2	2	00:45:00	10	10	00:45:00	°C		
	8d57c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	b4c27c050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	3f0198030000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	b5bbbf050000	True	5	2	2	00:45:00	10	10	00:45:00	°C		
	58482c050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		

Installation report

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
	d818bf050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	cfd5be050000	True	5	0	0	00:45:00	10	10	00:45:00	°C		
	26e83e050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	48e0bf050000	False	5	1	1	01:00:00	10	10	01:00:00	°C		
	84a8c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	0958c0050000	True	5	2	0	01:00:00	8	10	01:00:00	°C		
	5ecd3e050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	64a6bf050000	False	5	2	2	00:45:00	12	12	00:45:00	°C		
	7eeb97030000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	25f1be050000	False	5	2	2	00:45:00	10	10	00:45:00	°C		
	dda7c0050000	True	5	0	0	01:00:00	16	17	01:00:00	°C		
	5ea4bf050000	False	5	2	2	00:45:00	10	10	00:45:00	°C		
	842a98030000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	b958c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	bbc2c0050000	True	5	2	2	00:45:00	12	12	00:45:00	°C		
	dd59c0050000	True	5	2	2	00:45:00	10	10	00:45:00	°C		
	9f66c0050000	True	5	2	2	00:45:00	10	10	00:45:00	°C		

Installation report

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
	3b0cbf050000	False	5	0.88	0.88	00:45:00	10	10	00:45:00	°C		
	d95c3e050000	True	5	2	0	01:00:00	8	10	01:00:00	°C		
	cd8fbf050000	True	5	2	0	01:00:00	8	10	01:00:00	°C		
	aa1198030000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	10f63e050000	True	5	2	0	01:00:00	12	15	01:00:00	°C		
	f1c43e050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	baa8c0050000	True	5	2	2	00:45:00	8	9	00:45:00	°C		
	737b3e050000	True	5	2	0	01:00:00	8	10	01:00:00	°C		
	e8f6ec040000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	fef097030000	False	5	2	2	00:00:00	8	8	00:45:00	°C		
	1d354d030000	False	5	2	2	00:00:00	8	8	00:45:00	°C		
	a61d98030000	False	5	2	2	00:00:00	8	8	00:45:00	°C		
	122898030000	False	5	2	2	00:00:00	8	8	00:45:00	°C		
	f924bf050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	f795c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	84e0c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		

Installation report

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
S	6b5dce050000	True	20	15	15	00:45:00	30	30	00:45:00	°C		
	142c98030000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	f90e98030000	False	5	2	2	00:45:00	8	8	00:45:00	°C		
	b01298030000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	52d34d030000	False	5	2	2	00:45:00	8	8	00:45:00	°C		
	4a0d98030000	False	5	2	2	00:00:00	8	8	00:45:00	°C		
	330598030000	False	5	2	2	00:45:00	8	8	00:45:00	°C		
	131798030000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	85134d030000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	cab3bf050000	True	5	2	2	00:45:00	12	12	00:45:00	°C		
	1bfd97030000	False	5	2	2	00:00:00	8	8	00:45:00	°C		
	cf1998030000	False	5	2	2	00:00:00	8	8	00:45:00	°C		
	b0ef7c050000	False	5	2	2	00:00:00	8	8	00:45:00	°C		
	9bd0ce050000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	7094ce050000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	7a117d050000	False	5	2	2	00:45:00	8	8	00:45:00	°C		
	89052f030000	False	3	2	2	00:00:00	8	8	00:45:00	°C		

Installation report

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
	331098030000	False	5	2	2	00:00:00	8	8	00:45:00	°C		
	0e90bf050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	f8f397030000	False	5	2	2	00:00:00	8	9	00:45:00	°C		
	1ef697030000	True	5	2	2	00:00:00	8	10	00:45:00	°C		
	9c69c4040000	False	5	2	2	00:00:00	8	8	00:45:00	°C		
	20edbe050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	4c57c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	b65ac0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	f9d4bf050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	f0d5be050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	04b9bf050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	f32ec0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	eea4c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	a887c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	a9c7c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	9be47d050000	True	3	2	2	00:00:00	8	8	00:45:00	°C		
	23a11e050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	7e0198030000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	d35227060000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	1517bf050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		

Installation report

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
	0757c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	06d7bf050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	021598030000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	5e0ebf050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	1eaac0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	24214d030000	True	5	2	2	05:00:00	8	8	00:10:00	°C		
	510098030000	False	5	2	2	00:00:00	8	8	00:45:00	°C		
	d88a69030000	False	5	2	2	00:45:00	8	8	00:45:00	°C		
	83e0c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	721598030000	False	5	2	2	00:00:00	8	8	00:45:00	°C		
	7cf797030000	False	5	2	2	00:00:00	8	8	00:45:00	°C		
	8900bf050000	True	37	35	35	00:45:00	39	39	00:45:00	°C		
	b4bbbf050000	True	5	2	2	02:00:00	8	8	00:45:00	°C		
	6010ed040000	False	5	2	2	00:45:00	8	8	00:45:00	°C		
	11b53e050000	False	5	2	2	00:45:00	8	8	00:45:00	°C		
	6879bb020000	False	-20	-28	-28	00:45:00	-15	-15	00:45:00	°C		
	9839bb020000	False	-20	-26	-26	00:45:00	-14	-14	00:45:00	°C		
	8a653e050000	False	-20	-26	-26	00:45:00	-11	-11	00:45:00	°C		
	f1117d050000	False	-20	-25	-25	00:45:00	-15	-15	00:45:00	°C		

Installation report

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
	6a67c0050000	False	5	1	1	00:45:00	8	8	00:45:00	°C		
	4e5cc0050000	False	5	1	1	00:45:00	9	9	00:45:00	°C		
	60b7bf050000	False	5	2	2	00:45:00	8	8	00:45:00	°C		
	e4ab7d050000	False	-20	-27	-27	00:45:00	-13	-13	00:45:00	°C		
	e0a8c0050000	False	-20	-25	-25	00:45:00	-14	-14	00:45:00	°C		
	fec0ce050000	False	5	2	2	00:45:00	9	9	00:45:00	°C		
	705ece050000	False	4	1	1	00:45:00	8	8	00:45:00	°C		
	f4433f050000	False	4	2	2	00:45:00	8	8	00:45:00	°C		
	7cd77d050000	False	23	19	19	00:45:00	25	25	00:45:00	°C		
	2ce2be050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	1ad6bf050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	b3ab7d050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	6409bf050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	sns0_10294c300270	True	-80	-90	-90	00:45:00	-70	-70	00:45:00	°C		
	sns0_02294c300103	False	-80	-90	-90	00:45:00	-70	-70	00:45:00	°C		
	2551c0050000	False	5	2	2	00:45:00	8	8	00:45:00	°C		

Installation report

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
	d01cbf050000	False	-15	-25	-25	00:45:00	-10	-10	00:45:00	°C		
	d156c0050000	False	-15	-25	-25	00:45:00	-10	-10	00:45:00	°C		
	339341030000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	3d2c98030000	False	-35	-45	-45	00:45:00	-25	-25	00:45:00	°C		
	e8a741030000	True	-25	-38	-38	00:45:00	-15	-15	00:45:00	°C		
	4de897030000	True	-70	-80	-80	00:45:00	-60	-60	00:45:00	°C		
	642e0d040000	False	5	2	2	00:00:00	8	8	00:45:00	°C		
	cf6f3e050000	False	-15	-25	-25	00:45:00	-10	-10	00:45:00	°C		
	875c1e050000	False	-15	-25	-25	00:45:00	-10	-10	00:45:00	°C		
	02c63e050000	False	-15	-25	-25	00:45:00	-10	-10	00:45:00	°C		
	a4df3e050000	False	-20	-30	-30	00:45:00	-10	-10	00:45:00	°C		
	3994cd040000	False	7	2	2	00:30:00	8	8	00:45:00	°C		
	9b9d7d050000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	c50d7d050000	False	22	10	10	00:45:00	30	30	00:45:00	°C		
	bb9fc0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	4a3fce040000	False	23	19	19	00:45:00	25	25	00:45:00	°C		
	89052f030000	False	5	2	2	00:00:00	8	8	00:00:00	°C		
	e1ee7d050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	27403f050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		

Installation report

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
	15ef3e050000	True	-20	-30	-30	00:45:00	-10	-10	00:45:00	°C		
	136a3e050000	True	-20	-30	-30	00:45:00	-10	-10	00:45:00	°C		
	760598030000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	1479c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	sns_0_0b1e093001f7	False	1	0.5	0.5	00:05:00	1.5	1.5	00:05:00	230V		
	3f97ce050000	False	20	19	19	00:45:00	25	25	00:45:00	°C		
	a61d7d050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	96367d050000	False	22	17	17	00:45:00	25	25	00:45:00	°C		
	2dcd1f050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	31f03e050000	False	22	16	10	01:00:00	28	30	01:00:00	°C		
	480e3f050000	True	23	18	18	00:45:00	28	28	00:45:00	°C		
	1d97c0050000	False	-20	-40	-40	00:15:00	30	30	00:15:00	°C		
	sns0_02294c300146	False	-70	-90	-90	00:45:00	-40	-40	00:45:00	°C		
	715a3f050000	False	5	2	2	00:45:00	8	8	00:45:00	°C		
	42d1ce050000	False	20	18	18	00:45:00	23	23	00:45:00	°C		
	a86ac0050000	True	5	2	2	00:45:00	27	27	00:45:00	°C		

Installation report

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
	sns_0_11234d300235	True	50	15	15	00:45:00	75	75	00:45:00	%RH		
	2519bf050000	False	5	2	2	00:45:00	8	8	00:45:00	°C		
	f91b4d030000	True	5	2	2	00:00:00	8	8	00:45:00	°C		
	e2224d030000	False	-20	-30	-30	00:45:00	-10	-10	00:00:00	°C		
	sns_0_11234d300042	False	50	15	15	00:45:00	65	65	00:45:00	%RH		
	9942cb040000	False	20	10	10	00:45:00	30	30	00:45:00	°C		
	add97c050000	False	20	10	10	00:45:00	25	25	00:45:00	°C		
	sns_0_11234d300250	False	50	10	10	00:45:00	80	80	00:45:00	%RH		
	sns_0_0c1e0a3002ce	False	20	10	10	00:45:00	30	30	00:45:00	°C		
	c0d21d060000	False	-20	-40	-40	00:15:00	-4	-4	00:15:00	°C		
	90bf7d050000	True	20	18	18	00:45:00	27	27	00:45:00	°C		
	44c8c0050000	False	20	20	20	00:45:00	25	25	00:45:00	°C		
	1e831f050000	False	20	19	19	00:45:00	25	25	00:45:00	°C		
	627b3e050000	True	20	10	10	00:45:00	30	30	00:45:00	°C		
	8daa3e050000	True	20	18	18	00:45:00	26	26	00:45:00	°C		
	7fd47c050000	True	22	10	10	00:45:00	30	30	00:45:00	°C		
	31203f050000	True	22	15	15	00:45:00	30	30	00:45:00	°C		

Installation report

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
	88c93e050000	True	22	15	15	00:45:00	30	30	00:45:00	°C		
	sns_0_02234d300461	False	50	25	25	00:45:00	75	75	00:45:00	%rH		
	sns_0_02234d300458	False	50	25	25	01:01:00	75	75	01:01:00	%rH		
	01d4bf050000	False	5	1	1	00:45:00	10	10	00:45:00	°C		
	7a39c0050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	add77d050000	True	4	2	2	00:03:00	8	8	00:03:00	°C		
	96007d050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	1603bf050000	False	5	2	2	00:30:00	8	8	00:30:00	°C		
	b02a3f050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	e4223f050000	True	4	2	2	00:45:00	6	6	00:45:00	°C		
	5ab5bf050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	c3c63e050000	True	-18	-23	-23	00:45:00	-13	-13	00:45:00	°C		
	d23d3f050000	True	5	2	2	00:45:00	8	8	00:45:00	°C		
	e6cebe050000	True	-20	-25	-25	00:45:00	-15	-15	00:45:00	°C		
	66bf3e050000	True	-20	-25	-25	00:45:00	-15	-15	00:45:00	°C		
	sns0_02294c30010b	False	-80	-90	-90	00:45:00	-70	-70	00:45:00	°C		
	8e7ac0050000	True	-25	-30	-30	00:45:00	-20	-20	00:45:00	°C		
	187ac0050000	True	-25	-30	-30	00:45:00	-20	-20	00:45:00	°C		

Installation report

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
	d4d67d050000	True	-20	-25	-25	00:45:00	-15	-15	00:45:00	°C		
	2408bf050000	True	-20	-35	-35	01:00:00	-15	-15	01:00:00	°C		
	78563f050000	True	-25	-35	-40	01:00:00	-15	-14	01:00:00	°C		
	7eda7c050000	True	-20	-25	-25	00:45:00	-15	-15	00:45:00	°C		
	2ad5be050000	True	-30	-35	-35	00:45:00	-25	-25	00:45:00	°C		
	0bd73e050000	True	-25	-35	-40	00:45:00	-20	-16	00:45:00	°C		
	d6f63e050000	True	-20	-25	-25	00:45:00	-15	-15	00:45:00	°C		
	sns0_10294c30027b	True	-70	-80	-80	00:45:00	-60	-60	00:45:00	°C		
	bf9c3e050000	True	-20	-25	-25	00:45:00	-15	-15	00:45:00	°C		
	00a3bf050000	False	-25	-31	-32	00:45:00	-20	-20	00:45:00	°C		
	87593f050000	True	-20	-25	-30	01:00:00	-15	-10	01:00:00	°C		
	60c9c0050000	True	-20	-25	-25	00:45:00	-15	-15	00:45:00	°C		
	e01abf050000	True	-25	-29	-30	00:45:00	-21	-20	00:45:00	°C		
	67b23e050000	False	-25	-30	-30	00:45:00	-15	-10	00:45:00	°C		
	c4ef97030000	False	-22	-30	-30	00:01:00	-15	-15	00:01:00	°C		
	73d0bf050000	True	-20	-28	-28	00:45:00	-15	-15	00:45:00	°C		

Channel information

Channel Name	Sensor ID	Alarms activated	Set value	Low limit exceeded	Low time limit exceeded	Low time delay	High limit exceeded	High time limit exceeded	High time delay	Acquisition unit	Localization	Uncertainty
	96234d030000	False	-22	-24	-24	00:45:00	-20	-20	00:00:00	°C		
	ba300d040000	False	-22	-27	-27	00:45:00	-17	-17	00:00:00	°C		